


XA61 Sola

Side view



- (1) Z Axis (m)
 (2) Vertical (m)
 (3) Patient table

⊗: At this location, the force on a magnetically saturated ferromagnetic object is greatest.



The iso-magnetic contour lines are accurate to a value better than 1 %. There is very little influence from the environment to the scanner. The graphic shows the cover in nominal position. On a variety of scanners, the position can vary ± 5 mm in the axial direction. The maximum value therefore has a tolerance of ± 10 %, as small geometric deviations can cause a significant change in the value.